

A UWB imaging system to detect early breast cancer in heterogeneous breast phantom

Saleh Alshehri^a; Adznan Jantan^a; R.S.A. Raja Abdullah^a; Rozi Mahmud^b; Sabira Khatun^c; Zaiki Awang^d

^aDepartment of Computer and Communication systems Engineering, Faculty of Engineering,

^bDepartment of Imaging, Faculty of Medicine and Health Sciences, Universiti Putra Malaysia, Serdang, Malaysia

^cDepartment of Computer Systems & Networks, Faculty of Computer systems & Software Engineering, Universiti Malaysia Pahang, Pahang, Malaysia.

^dMicrowave Technology Center Faculty of Electrical Engineering Universiti Teknologi Mara, 40400, Shah Alam, Selangor, Malaysia

ABSTRACT

This paper presents an experimental early breast cancer detection system in terms of heterogeneous breast phantom. The system consists of commercial Ultrawide-Band (UWB) transceivers and our developed Neural Network (NN) based Pattern Recognition (PR) software for imaging. A simple way to construct cancer- tissue and heterogeneous breast phantom using available low cost materials and their mixtures is also proposed here. The materials are: (i) A mixture of petroleum jelly, soy oil, wheat flour and water as heterogeneous tissue; (ii) A particular glass as skin; and (iii) A specific mixture of water and wheat flour as cancer- tissue. All the materials and their mixtures are considered according to the ratio of the dielectric properties of the breast tissues. To experimentally detect cancer, the UWB signals are transmitted from one side of the breast phantom and received from opposite side diagonally. By using discrete cosine transform (DCT) of the received signals, a Neural Network (NN) is trained, tested and interfaced with the UWB transceiver to form the complete system. The achieved detection rate of cancer cell's existence, size and location are approximately 100%, 93.1% and 93.3% respectively.

KEYWORDS:

Early Breast Cancer Detection; Neural Network; Breast Phantom For UWB Imaging.

REFERENCES

1. Bindu, G., Lonappan, A., Thomas, V., Ananadan, C. K., Mathew, K.T.: Active microwave imaging for breast cancer detection. Progress In Electromagnetic Research, PIER, vol. 58, 149-169 (2006)
2. Lim, H. B. , Nhung, N. T., Li, E., Thang, N. D.: Confocal microwave imaging for breast cancer detection: delay-multiply-and-sum image reconstruction algorithm. IEEE Transaction on Biomedical Engineering, vol. 55, no. 6, 1697-1704 (2008)
3. Fear, E. C., Li, X., Hagness, S. C., Stuchly, M.A.: Confocal microwave imaging for breast tumor detection: localization of tumors in three dimensions. IEEE Transactions on Biomedical Engineering, vol. 49, no. 8, 812-822 (2002)
4. Fear, E. C., Still, J., Stuchly, M.A.: Experimental feasibility study of confocal microwave imaging for breast tumor detection. IEEE Transactions on Microwave Theory and Techniques, vol. 51, no. 3, 887- 897 (2003)
5. Li, X., Davis, S. K., Hagness, S. C., Weide, D. W., Veen, B. D.: Microwave imaging via space-time beam forming: experimental investigation of tumor detection in multilayer breast phantoms. IEEE Trans. Microwav. Theory Techniques, vol. 52, no. 8, 1856-1865, (2004)