

# Ultra wide Band (UWB) Based Early Breast Cancer Detection using Artificial Intelligence

Bifta Sama Bari<sup>1</sup>, Sabira Khatun<sup>1</sup>, Kamarul Hawari Ghazali<sup>1</sup>, Md. Moslemuddin Fakir<sup>2</sup>, Wan Nur Azhani W. Samsudin<sup>1</sup>, Mohd Falfazli Mat Jusof<sup>1</sup>, Mamunur Rashid<sup>1</sup>, Minarul Islam<sup>1</sup>, and Mohd Zamri Ibrahim<sup>1</sup>

<sup>1</sup>Faculty of Electrical and Electronics engineering, Universiti Malaysia Pahang, Malaysia

<sup>2</sup>CARIFF, Faculty of Chemical & Natural Resources Engineering, Universiti Malaysia Pahang, Malaysia

biftasama\_120@yahoo.com      sabirakhatun@ump.edu.my  
kamarul@ump.edu.my          moslemuddin@ump.edu.my  
nurazhani@ump.edu.my        mfalfazli@ump.edu.my  
mamun110218@gmail.com       sumankhancsejust@gmail.com  
zamri@ump.edu.my

**Abstract.** Breast cancer is a silent killer malady among women community all over the world. The death rate is increased as it has no syndrome at early stage. There is no remedy; hence, detection at the early stage is crucial. Usually, women do not go to clinic/hospital for regular breast health checkup unless they are sick. This is due to long queue and waiting time in hospital, high cost, people's busy schedule, and so on. Recently, several research works has been done on early breast cancer detection using Ultra Wide Band (UWB) technology because of its non-invasive and health-friendly nature. Each proposed UWB system has its own limitation including system complexity, expensive, expert operable in clinic. To overcome these problems, a system is required which should be simple, cost-effective and user-friendly. This chapter presents the development of a user friendly and affordable UWB system for early breast cancer detection utilizing Artificial Neural Network (ANN). A feed-forward back propagation Neural Network (NN) with 'feedforwardnet' function is utilized to detect the cancer existence, size as well as location in 3-dimension (3D). The hardware incorporates UWB transceiver and a pair of pyramidal shaped patch antenna to transmit and receive the UWB signals. The extracted features from the received signals have been fed into the NN module to train, validate, and test. The average system's performance efficiency in terms of tumor/cancer existence, size and location are approximately 100%, 92.43% and 91.31 % respectively. Here, in our system, use of 'feedforwardnet' function; detection-combination of tumor/cancer existence, size and location in 3D along with improved performance is a new addition compared to other related researches and/or existing systems. This may become as a promising user-friendly system in near future for early breast cancer detection in domestic environment with low cost and to save precious life.

20. Vijayasarveswari, V., Jusoh, M., Khatun, S.: Experimental UWB based Efficient Breast Cancer Early Detection. *Indian Journal of Science and Technology*, 10(12), 1-6. (2017).
21. Vijayasarveswari, V., Jusoh, M., Sabapathy, T., Aliana1, R., Khatun, S., Ahmad, Z.A., Osman, M.N.: Performance Verification on UWB Antennas for Breast Cancer Detection. *MATEC Web of Conferences*, vol. 140, pp. 01004 (2017).
22. Meaney, P.M., Paulsen, K. D., Hartov, A., Crane, R. K.: An Active Microwave Imaging System for Reconstruction of 2-D Electrical Property Distributions. *IEEE Trans on Biomed Eng.* 42(10),1017–1026 (1995).
23. Fear, E.C., Sill, J.M.: Preliminary Investigations of Tissue Sensing Adaptive Radar for Breast Tumor Detection. In: *Proceedings of the 25th Annual International Conf of the IEEE Eng in Med and Biology Society*, vol. 4, pp. 3787–3790, Mexico, (2003).
24. Abu Bakar, A., Ireland, D., Abbosh, A. M., Wang, Y.: Experimental Assessment of Microwave Diagnostic Tool for Ultra-Wideband Breast Cancer Detection. *Prog in Electromagnetics Research M*, vol. 23, 109-121 (2012).
25. Shahzad, A., O'Halloran, M., Jones, E., Glavin, M.: A multistage selective weighting method for improved microwave breast tomography. *Computerized Med Imaging and Graphics*, vol. 54, 6-15 (2016).
26. Salleh, S.H.M., Othman, M.A., Ali, N., Sulaiman, H.A., Misran, M.H., Abd. Aziz, M.Z.A.: Microwave Imaging Technique Using UWB Signal for Breast Cancer Detection. *ARNP J. of Eng and App Sci.* 10(2), 723-727 (2015).
27. RCM Pulse-On UWB devices. Time Domain Corporation, Comings Research Park; 330 Wynn Drive, Suite 300, Huntsville, AL 35805, USA, (2017).
28. Lazebnik, M., Madsen, E.L., Frank, G.R., Hagness, S.C.: Tissue-mimicking phantom materials for narrowband and ultrawideband microwave applications. *Phys. Med. Biol.* 50: 4245-4258 (2005).
29. Porter, E., Fakhoury, J., Oprisor, R., Coates, M., Popović, M.: Improved tissue phantoms for experimental validation of microwave breast cancer detection. *Antennas and Propagation (EuCAP)*, pp. 1-5. April (2010).
30. Conceicao, R. C., O'Halloran, M., Jones, E., Glavin, M.: Investigation of Classifiers for Early-stage Breast Cancer Based on Radar Target Signatures. *Progress In Electromagnetics Research*, 105, 295–311 (2010).
31. Santorelli, A., Porter, E., Kirshin, E., Liu, Y.J., Popović, M.: Investigation of Classifiers for Tumor Detection with an Experimental Time-Domain Breast Screening System. *Progress In Electromagnetics Research*. 144, 45-57 (2014).