

Thermal Oxidation Promotes Growth of Nanocrystalline Diamond on Biomedical Grade Co-Cr-Mo Alloy

H. Mas Ayu, S. Izman, R. Daud, A. Shah, S. H. Tomadi, M. S. Dambatta

1. Faculty of Mechanical Engineering and Manufacturing, Universiti Malaysia Pahang, 26600 Pekan, Pahang, Malaysia.
2. Faculty of Mechanical Engineering, Universiti Teknologi Malaysia, 81310 Skudai, Johor, Malaysia.
3. Faculty of Technical and Vocational, Universiti Pendidikan Sultan Idris, 35900 Tanjong Malim, Perak, Malaysia
4. Department of Manufacturing and Materials Engineering, International Islamic University Malaysia, P.O. Box 10, 50728 Kuala Lumpur, Malaysia
5. Department of Mechanical Engineering, Faculty of Engineering, Kano University of Science and Technology Wudil, P. M. B 3244, Kano State, Nigeria.

Abstract:

Diamond coatings are employed to yield significant benefits in applications such as for cutting tools, optical lenses, biomedical components, microelectronics, engineering and thermal management systems. Although there are many research reporting the successful of diamond coating on titanium, tungsten carbide and steel alloys but there are still lacking of research on the cobalt based alloy as the substrate. In order to coat diamond on these metals substrate, chemical vapor deposition (CVD) technique is commonly used. This paper reports on investigations of nano-crystalline diamond (NCD) coating on different carbon content of cobalt based alloys. Emphasis is given to achieve good adhesion of NCD coating on high and low carbon content of cobalt-chromium-molybdenum (Co-Cr-Mo) alloys using two different processes of surface pretreatment such as mechanical roughening the sample surface ($\approx 0.3\mu\text{m}$) and by using thermal oxidation to create oxide interlayer. The results revealed that most of the coating on samples was peel-off on roughened surface. However, there were small portion of NCD coating that still intact on sample with oxide interlayer. The thickness NCD coating obtained was approximately $5\mu\text{m}$. Since the adhesion strength of the diamond coatings were very poor and easily delaminated, scratch test could not be performed on both sample conditions. Surface morphology and characterization of diamond coatings were investigated by Scanning Electron Microscopy (SEM) and X-ray diffraction respectively.

Keywords: Co-Cr-Mo Alloy; Thin Film; NCD Coating; Surface Morphology; Biomaterial

Acknowledgments

This research project was funded under the Fundamental Research Grant Scheme RDU190130 and research grant RDU1703319 which provided by Ministry of Higher Education Malaysia and Faculty of Mechanical Engineering, Universiti Malaysia Pahang respectively. The authors also would like to acknowledge Faculty of Mechanical Engineering, Universiti Teknologi Malaysia for providing their facilities in conducting this research.